

2.9. PHARMACEUTICAL TECHNICAL PROCEDURES

01/2005:20901

2.9.1. DISINTEGRATION OF TABLETS AND CAPSULES

The disintegration test determines whether tablets or capsules disintegrate within the prescribed time when placed in a liquid medium in the experimental conditions prescribed below.

Disintegration is considered to be achieved when:

- a) no residue remains on the screen, or
- b) if there is a residue, it consists of a soft mass having no palpably firm, unmoistened core, or
- c) only fragments of coating (tablets) or only fragments of shell (capsules) remain on the screen; if a disc has been used (capsules), fragments of shell may adhere to the lower surface of the disc.

Use apparatus A for tablets and capsules that are not greater than 18 mm long. For larger tablets or capsules use apparatus B.

TEST A - TABLETS AND CAPSULES OF NORMAL SIZE

Apparatus. The main part of the apparatus (Figure 2.9.1-1) is a rigid basket-rack assembly supporting 6 cylindrical transparent tubes 77.5 ± 2.5 mm long, 21.5 mm in internal diameter, and with a wall thickness of about 2 mm. Each tube is provided with a cylindrical disc 20.7 ± 0.15 mm in diameter and 9.5 ± 0.15 mm thick, made of transparent plastic with a relative density of 1.18 to 1.20 or weighing 3.0 ± 0.2 g. Each disc is pierced by 5 holes 2 mm in diameter, 1 in the centre and the other 4 spaced equally on a circle of radius 6 mm from the centre of the disc. On the lateral surface of the disc, 4 equally spaced grooves are cut in such a way that at the upper surface of the disc they are 9.5 mm wide and 2.55 mm deep and at the lower surface 1.6 mm square. The tubes are held vertically by 2 separate and superimposed rigid plastic plates 90 mm in diameter and 6 mm thick with 6 holes. The holes are equidistant from the centre of the plate and equally spaced. Attached to the under side of the lower plate is a piece of woven gauze made from stainless steel wire 0.635 mm in diameter and having mesh apertures of 2.00 mm. The plates are held rigidly in position and 77.5 mm apart by vertical metal rods at the periphery, a metal rod is also fixed to the centre of the upper plate to enable the assembly to be attached to a mechanical device

2. Methods of analysis




Figure 2.9.1-1. - Apparatus A
Dimensions in millimetres

capable of raising and lowering it smoothly at a constant frequency between 29 and 32 cycles per minute, through a distance of 50 mm to 60 mm.

The assembly is suspended in the specified liquid in a suitable vessel, preferably a 1 litre beaker. The volume of the liquid is such that when the assembly is in the highest position the wire mesh is at least 15 mm below the surface of the liquid, and when the assembly is in the lowest position the wire mesh is at least 25 mm above the bottom of the beaker and the upper open ends of the tubes remain above the surface of the liquid. A suitable device maintains the temperature of the liquid at 35-39 °C.

The design of the basket-rack assembly may be varied provided the specifications for the tubes and wire mesh are maintained.

Method. In each of the 6 tubes, place one tablet or capsule and, if prescribed, add a disc; suspend the assembly in the beaker containing the specified liquid. Operate the apparatus for the prescribed period, withdraw the assembly and examine the state of the tablets or capsules. To pass the test, all the tablets or capsules must have disintegrated.

TEST B – LARGE TABLETS AND LARGE CAPSULES

Apparatus. The main part of the apparatus (Figure 2.9.1-2) is a rigid basket-rack assembly supporting 3 cylindrical transparent tubes 77.5 ± 2.5 mm long, $33.0 \text{ mm} \pm 0.5$ mm in internal diameter, and with a wall thickness of 2.5 ± 0.5 mm. Each tube is provided with a cylindrical disc 31.4 ± 0.13 mm in diameter and 15.3 ± 0.15 mm thick, made of transparent plastic with a relative density of 1.18 to 1.20 or weighing 13.0 ± 0.2 g. Each disc is pierced by 7 holes, each 3.15 ± 0.1 mm in diameter, 1 in the centre and the other 6 spaced equally on a circle of radius 4.2 mm from the centre of the disc. The tubes are held vertically by 2 separate and superimposed rigid plastic plates 97 mm in diameter and 9 mm thick, with 3 holes. The holes are equidistant from the centre of the plate and equally spaced. Attached to the under side of the lower plate is a piece of woven gauze made from stainless steel wire 0.63 ± 0.03 mm in diameter and having mesh apertures of 2.0 ± 0.2 mm. The plates are held rigidly in position and 77.5 mm apart by vertical metal rods at the periphery, a metal rod is also fixed to the centre of the upper plate to enable the assembly to be attached




Figure 2.9.1-2. – Apparatus B
Dimensions in millimetres

to a mechanical device capable of raising and lowering it smoothly at constant frequency between 29 and 32 cycles per minute, through a distance of 55 ± 2 mm.

The assembly is suspended in the specified liquid medium in a suitable vessel, preferably a 1 litre beaker. The volume of the liquid is such that when the assembly is in the highest position the wire mesh is at least 15 mm below the surface of the liquid, and when the assembly is in the lowest position the wire mesh is at least 25 mm above the bottom of the beaker and the upper open ends of the tubes remain above the surface of the liquid. A suitable device maintains the temperature of the liquid at $35\text{--}39$ °C.

The design of the basket-rack assembly may be varied provided the specifications for the tubes and wire mesh are maintained.

Method. Test 6 tablets or capsules either by using 2 basket-rack assemblies in parallel or by repeating the procedure. In each of the 3 tubes, place one tablet or capsule and, if prescribed, add a disc; suspend the assembly in the beaker containing the specified liquid. Operate the apparatus for the prescribed period, withdraw the assembly and examine the state of the tablets or capsules. To pass the test, all 6 of the tablets or capsules must have disintegrated.

01/2005:20902

2.9.2. DISINTEGRATION OF SUPPOSITORIES AND PESSARIES

The disintegration test determines whether the suppositories or pessaries soften or disintegrate within the prescribed time when placed in a liquid medium in the experimental conditions described below.

Disintegration is considered to be achieved when:

- dissolution is complete,
- the components of the suppository or pessary have separated: melted fatty substances collect on the surface of the liquid, insoluble powders fall to the bottom and soluble components dissolve, depending on the type of preparation, the components may be distributed in one or more of these ways,
- there is softening of the sample that may be accompanied by appreciable change of shape without complete separation of the components, the softening is such that the suppository or pessary no longer has a solid core offering resistance to pressure of a glass rod,
- rupture of the gelatin shell of rectal or vaginal capsules occurs allowing release of the contents,
- no residue remains on the perforated disc or if a residue remains, it consists only of a soft or frothy mass having no solid core offering resistance to pressure of a glass rod (vaginal tablets).

Apparatus. The apparatus (Figure 2.9.2.-1) consists of a sleeve of glass or suitable transparent plastic, of appropriate thickness, to the interior of which is attached by means of three hooks a metal device consisting of two perforated stainless steel discs each containing 39 holes 4 mm in diameter; the diameter of the discs is similar to that of the interior of the sleeve; the discs are about 30 mm apart. The test is carried out using three such apparatuses each containing a single sample. Each apparatus is placed in a beaker with a capacity of at least 4 litres filled with water

maintained at $36\text{--}37$ °C, unless otherwise prescribed. The apparatuses may also be placed together in a vessel with a capacity of at least 12 litres. The beaker is fitted with a slow stirrer and a device that will hold the cylinders vertically not less than 90 mm below the surface of the water and allow them to be inverted without emerging from the water.

Method. Use three suppositories or pessaries. Place each one on the lower disc of a device, place the latter in the sleeve and secure. Invert the apparatuses every 10 min. Examine the samples after the period prescribed in the monograph. To pass the test all the samples must have disintegrated.




Figure 2.9.2.-1. – Apparatus for disintegration of suppositories and pessaries

Dimensions in millimetres

METHOD OF OPERATION FOR VAGINAL TABLETS

Use the apparatus described above, arranged so as to rest on the hooks (see Figure 2.9.2.-2). Place it in a beaker of suitable diameter containing water maintained at $36\text{--}37$ °C with the level just below the upper perforated disc. Using a pipette, adjust the level with water at $36\text{--}37$ °C until a uniform film covers the perforations of the disc. Use three vaginal tablets. Place each one on the upper plate of an apparatus and cover the latter with a glass plate to maintain appropriate conditions of humidity. Examine the state of the samples after the period prescribed in the monograph. To pass the test all the samples must have disintegrated.